

Influence of beam pitch and tube current on lung nodule 64- MDCT - chest protocol : A phantom study

Chatnapa Nuntue*

Anchali Krisanachinda**

Nuntue C, Krisanachinda A. Influence of beam pitch and tube current on lung nodule 64- MDCT - chest protocol: A phantom study. Chula Med J 2016 Nov – Dec;60(6): 617 - 28

Background : *Multi-Detector Computed Tomography (MDCT) is a powerful modality to detect small pulmonary nodules while the benefits of CT exceed the harmful effects of radiation exposure in patients; thus, CT doses should be kept as low as reasonably achievable. The way to reduce radiation dose is based on modification of scanning parameters.*

Objective : *The purpose of this study is to investigate the radiation dose and image quality affected by beam pitch and tube current for chest CT scan using a lung man phantom with lung nodules.*

Designs : *Experimental study (in vitro).*

Setting : *Department of Radiology, King Chulalongkorn Memorial hospital, Bangkok.*

Material and Method : *A lung man chest phantom with various simulated nodules was scanned by 64-MDCT with standard and high beam pitch technique and varying tube current (40-300 mA). The quantitative image quality was investigated by the Contrast to Noise Ratio (CNR). The qualitative image quality was assessed by two reviewers, using five -point scale for nodules detectability. The volume CT dose index ($CTDI_{vol}$) were recorded to evaluate the radiation dose for standard beam pitch technique and high beam pitch technique of lung nodules CT scan.*

* Department of Radiological Technology, Faculty of Allied Health Science, Thammasat University

** Department of Radiology, Faculty of Medicine, Chulalongkorn University

Results : *When the tube current, mA, was reduced from 300 to 40, the CTDI_{vol} decreased from 12.3 to 1.60 mGy for standard beam pitch technique and from 8.60 to 1.18 for high beam pitch technique. Also, comparing the standard beam pitch technique and high beam pitch technique on various tube current (mA), the average radiation dose reduced by 30%. The average CNR decreased when high beam pitch was used. The scoring on image quality by two reviewers was in good agreement.*

Conclusions : *Beam pitch and tube current is an important factor for CT chest image quality. The high beam pitch CT showed a low radiation dose and short scan time when compared to the standard beam pitch with diagnostic image quality maintained.*

Keywords : *Chest scan, beam pitch, 64-MDCT.*

Correspondence to: Krisanachinda A. Department of Radiology, Faculty of Medicine,
Chulalongkorn University, Bangkok 10330, Thailand.

Received for publication. August 4, 2016.

ฉัตรนภา นันต์ดี, อัญชลี กฤษณจินดา. อิทธิพลของค่าบีมพิตซ์และกระแสหลอดต่อก้อนเนื้อ
ในปอด สำหรับโปรโตคอลทรวงอกด้วยเอกซเรย์คอมพิวเตอร์ 64 หัววัด: การศึกษาในหุ่น
จำลอง. จุฬาลงกรณ์เวชสาร 2559 พ.ย. – ธ.ค.;60(6): 617 – 28

- เหตุผลการทำวิจัย** : เครื่องเอกซเรย์คอมพิวเตอร์ชนิดหลายหัววัดเป็นเครื่องมือที่มีบทบาท
ในการตรวจพบก้อนเนื้อขนาดเล็กในปอด ในขณะที่ประโยชน์ที่ได้รับจาก
การตรวจด้วยเครื่อง เอกซเรย์คอมพิวเตอร์มีค่ามากกว่าปริมาณรังสีที่ผู้ป่วย
ได้รับ ดังนั้นปริมาณรังสีที่ผู้ป่วยได้รับควรเป็นไปตามหลักสากลโดยใช้ปริมาณ
รังสีน้อยที่สุดและเกิดประโยชน์สูงสุด วิธีลดปริมาณรังสีขึ้นกับการปรับ
เปลี่ยนพารามิเตอร์ของการตรวจ
- วัตถุประสงค์** : เพื่อประเมินปริมาณรังสี และคุณภาพของภาพ เมื่อเปลี่ยนค่ากระแสหลอด
และค่าบีมพิตซ์ของเครื่องเอกซเรย์คอมพิวเตอร์ในการตรวจช่องอก โดยใช้
หุ่นจำลองส่วนช่องอกและก้อนเนื้อในปอด
- รูปแบบการวิจัย** : การศึกษาเชิงทดลอง (นอกร่าง)
- สถานที่ทำการศึกษา** : ฝ่ายรังสีวิทยา โรงพยาบาลจุฬาลงกรณ์ สภากาชาดไทย
- วิธีการศึกษา** : ทำการสแกนหุ่นจำลองส่วนช่องอกพร้อมทั้งก้อนเนื้อในปอดโดยเครื่อง
เอกซเรย์คอมพิวเตอร์ 64 หัววัด ด้วยค่าบีมพิตซ์มาตรฐานและค่าบีมพิตซ์
สูงขึ้น และทำการตั้งค่ากระแสหลอด ตั้งแต่ 40 ถึง 300 มิลลิแอมแปร์
ประเมินคุณภาพของภาพเชิงปริมาณโดยการหาอัตราส่วนระหว่างความ
คมชัดและสัญญาณรบกวนคุณภาพของภาพ ในเชิงคุณภาพประเมิน
โดยผู้ประเมิน 2 คนที่มีประสบการณ์ โดยใช้เกณฑ์การประเมินการตรวจพบ
ก้อนเนื้อในปอดด้วยคะแนน 1- 5 และบันทึกค่าปริมาณรังสีเชิงปริมาตร
ซีทีดีไอวอลลุ่ม (มิลลิเกรย์) ที่อ่านได้จากหน้าจอสำหรับค่าบีมพิตซ์มาตรฐาน
และค่าบีมพิตซ์ที่สูงขึ้นของการตรวจสแกนส่วนช่องอกและก้อนเนื้อในปอด
- ผลการศึกษา** : เมื่อลดค่ากระแสหลอดจาก 300 ถึง 40 มิลลิแอมแปร์ มีผลทำให้ค่าปริมาณ
รังสีเชิงปริมาตรซีทีดีไอวอลลุ่ม ลดลงจาก 12.3 เป็น 1.60 มิลลิเกรย์
สำหรับค่าบีมพิตซ์มาตรฐาน และลดลงจาก 8.60 เป็น 1.18 มิลลิเกรย์
สำหรับค่าบีมพิตซ์ที่สูงขึ้น และเมื่อเปรียบเทียบระหว่างค่าบีมพิตซ์มาตรฐาน
และค่าบีมพิตซ์ที่สูงขึ้น ในค่ากระแสหลอดต่าง ๆ มีผลให้ปริมาณรังสีลดลง
เฉลี่ยร้อยละ 30 เปร้ออัตราส่วนระหว่างความคมชัดและสัญญาณรบกวน
คุณภาพของภาพมีค่าลดลงเมื่อใช้ค่าบีมพิตซ์ที่สูงขึ้น คะแนนการประเมิน
คุณภาพของภาพโดยผู้ประเมิน 2 คนมีความสอดคล้องกันดี

- สรุป** : บี้มพิตซ์และกระแสหลอด เป็นพารามิเตอร์ที่มีความสำคัญต่อคุณภาพของภาพในการตรวจเอกซเรย์คอมพิวเตอร์ส่วนช่องอก ผลกระทบของค่าพิตซ์ที่สูงขึ้นแสดงให้เห็นว่าปริมาณรังสี และเวลาในการสแกนลดลง เมื่อเทียบกับค่าบี้มพิตซ์มาตรฐานในขณะที่คุณภาพของภาพเพียงพอต่อการวินิจฉัย
- คำสำคัญ** : การตรวจเอกซเรย์คอมพิวเตอร์ช่องอก, บี้มพิตซ์, เครื่องเอกซเรย์คอมพิวเตอร์ชนิด 64 หัววัด.

At present, the number of Multi-Detector Computed Tomography (MDCT) chest examination increases for clinical diagnostic especially in lung cancer, patient checked up for lung screening and patient with small solitary pulmonary nodule that required follow up regarding the size of lesion several times for malignancy investigation. ^(1 - 2) This is a major concern because of the associated radiation exposure is potential linked to the pathogenesis of cancer. The CT doses seem to be lower in updated reports, as the concerns for radiation and the advances in CT technology. ⁽³⁾ However, CT doses should be kept as low as reasonably achievable. During the last decade, CT dose reduction strategies have been realized by using various techniques such as tube voltage reduction, scan length optimization, individualization of scan protocol and utilization of automatic exposure control ^(4 - 6) while diagnostic capability can be maintained.

One of the important technical parameter to reduce the radiation dose is the beam pitch, the ratio between the table move per rotation time and the collimator width. The beam pitch of 1.0 facilitates an

acquisition with no overlap or gap, while less than 1.0 facilitates an overlapping acquisition and higher patient dose, and greater than 1.0 facilitates a gap acquisition with low patient dose and image quality. ⁽⁷⁾ The purpose of this study is to investigate the radiation dose and image quality when various tube currents and beam pitch CT had been applied for chest scans using a lung man chest phantom with lung nodules.

Materials and Methods

Lung man chest phantom study

The study was performed at the Department of Radiology, King Chulalongkorn Memorial Hospital. The lung man chest phantom manufactured by Kyoto Kagaku Co.Ltd., Japan (Fig.1) has x-ray attenuation relatively to the human tissues. In order to mimic pulmonary lesions, various simulated nodule of spheres 3,5,8,10 and 12 mm in diameter (Fig.2) with CT number of 100 HU located in the lung field of the lung man chest phantom was scanned by 64-MDCT(Brilliance 64, Philips Health Care).



Figure 1. External (left) and internal (right) appearances of Lung man chest phantom.



Figure 2. Simulated nodule of spheres 3, 5, 8, 10 and 12 mm in diameter.

The reference value with the following protocol were a tube voltage of 120 kVp, tube current of 300 mA, helical scan mode, 64×0.625 mm detector configuration, 5 mm slice thickness, reconstruction slice thickness 1 mm on lung window with a center of -600 HU and width of 1600 HU rotation time of 0.5 sec, Scan FOV 350 mm, 350 mm scan length and standard filter. The examinations of lung man chest phantom were scanned from the apex of the lung to the lower costal margin by varying standard beam pitch (0.798) and high beam pitch (1.141) at tube current 40, 60, 80, 100, 200 and 300 mA while the other parameters were fixed. The scan time on lung man chest phantom examinations with standard and high beam pitch was 5.8 and 4.2 second respectively. In order to assess the radiation dose on lung man chest phantom, The Volume Computed Tomography Dose Index ($CTDI_{vol}$, mGy) were recorded from CT monitor.

The $CTDI_{vol}$ has been verified for the accuracy, reproducibility and reliable before collecting the data. The polymethylmethacrylate body phantom (32 cm diameter) was used. The

measurement on $CTDI_{vol}$ was performed by placing a 100 mm pencil-shaped ionization chamber at the center and the peripheral positions of the body phantom at the isocenter of the CT bore by following to the IAEA Human Health No.19 protocol.⁽⁸⁾ The $CTDI_{vol}$ displayed on CT monitor were recorded and compared with the measured values in percentage of difference.

Quantitative image quality

The image quality was determined by the Contrast to Noise Ratio (CNR) which measured by placing the 2 circular ROIs of similar area within the nodule and background at the same slice (Fig.3).

The CT number within the ROIs will be recorded in order to calculate the CNR by the following equation:

$$CNR = \frac{(CT_n - CT_{bg})}{SD_{bg}}$$

Where CT_n is the CT number of nodule, CT_{bg} is the CT number of background and SD_{bg} is standard deviation of background.




Figure 3. The location of ROIs for measuring the CT number of nodule and background.

Qualitative image quality

The image quality was assessed by two reviewers who have similar experience. They were blinded random order to the CT scanning parameter. The reviewers independently scored the image for nodule detectability on the same PACS monitor and under the same environment by using a five point scale: Score 5 = Excellent (visualize all simulated nodules with sharp edge); 4 = Good (visualize clearly 5 mm, partly visualize 3 mm in diameter of simulated nodule); 3 = Acceptable (visualize clearly 10 and 8 mm, partly visualize 5 mm in diameter of simulated nodule); 2 = Poor (visualize clearly 12 mm, partly visualize 10 mm in diameter); 1 = Unacceptable (visualize blur of all simulated pulmonary nodules).

Statistical analysis

In order to evaluate inter-observer reliability between two reviewers, weighted kappa were used for qualitative image analysis with $k < 0.20$, poor agreement; $k = 0.20 - 0.40$, fair agreement; $k = 0.41 - 0.60$, moderate agreement; $k = 0.61 - 0.80$, good

agreement; and $k = 0.81 - 1.0$, almost perfect agreement.⁽⁹⁾

Results

For the CTDI_{vol} verification, the percent difference of measured values and the monitor displayed of the polymethylmethacrylate body phantom (32 cm) were in tolerance range according to IAEA Human Health No.19.⁽⁸⁾

The results of the radiation dose at various tube current (mA) with standard beam pitch technique and high beam pitch technique are illustrated in Table 1. When the tube current from 300 to 40 mA resulted in the CTDI_{vol} decreased from 12.3 to 1.60 mGy for standard beam pitch technique and from 8.60 to 1.18 mGy for high beam pitch technique. The study showed that the lower CTDI_{vol} of high beam pitch technique than the standard beam pitch technique and the average percentage of CTDI_{vol} reduction is about 30% when comparing between standard beam pitch technique and high beam pitch technique on various tube current (mA).

Table 1. Radiation exposure from different tube current (mA), beam pitch and %CTDI_{vol} reduction.

Tube current (mA)	Beam pitch	CTDI _{vol} (mGy)	%CTDI _{vol} reduction
40	0.798	1.60	26.25%
	1.141	1.18	
60	0.798	2.49	31.73%
	1.141	1.70	
80	0.798	3.27	29.66%
	1.141	2.30	
100	0.798	4.12	32.04%
	1.141	2.80	
200	0.798	8.18	31.54%
	1.141	5.60	
300	0.798	12.30	30.08%
	1.141	8.60	

For quantitative image quality, the bar chart representing the average CNR value obtained from all simulated nodule size plotted against with the tube current (mA) for standard beam pitch technique and high beam pitch technique is illustrated in Figure 4. It is demonstrated that the CNR decrease when beam pitch value increase because the table moves faster at higher beam pitch, thus less photon reaching the detectors. The example of measuring image for calculated CNR of tube current 60 mA compare between standard beam pitch technique and high

beam pitch technique on nodule 8 mm in diameter is illustrated in Figure 5.

For qualitative image quality, Figure 6 showed the chest image for visualize of nodule detectability, for example chest image with nodule 10 mm in diameter for 300 mA with standard beam pitch (Fig. 6A), the chest image with nodule 10 mm in diameter for 300 mA with high beam pitch (Fig. 6B) and nodule 8 mm in diameter for 200 mA with high beam pitch (Fig.6C)




Figure 4. The relationship between CNR, tube current at different beam pitch techniques.




Figure 5. The measuring image for calculated CNR compare between standard beam pitch technique (left) and high beam pitch technique (right) on nodule 8 mm in diameter.




Figure 6 (A)

Figure 6 (B)

Figure 6 (C)

Figure 6. The example of chest image for visualize of nodule detectability.

The agreement of image quality scored by two reviewers was determined by calculating weighted Kappa of the variation tube current (mA) and beam pitch. The two reviewers had substantially inter-observer agreement for nodule detectability as illustrated in Table 2.

The k-value from weighted kappa is usually interpreted the strength of agreement, and the k-value of 0.735 is obtained from this study, which means the strength of agreement is good.

Discussion

With technological advances of computed tomography (CT), the modern MDCT provides a useful function that the user can adjust parameters to modify

the image quality relate to the radiation dose.⁽¹⁰⁾ One of the important parameter which affects to the image quality is beam pitch. In this study, the high beam pitch technique on Chest protocol has been applied to determine the efficiency of the radiation dose reduction and maintain image quality for the Chest protocol.

For radiation dose, the CTDI_{vol} reduced around 30% when compared between high beam pitch technique and standard beam pitch technique. The CTDI_{vol} of high beam pitch technique (1.141) is lower than the standard beam pitch technique (0.798) in all various tube current (mA) because in theory, increase beam pitch, which the table move fast link to the shortening total scan time and consequently affect

Table 2. The inter-observer agreement for nodule detectability by two reviewers.

Reviewer B	Reviewer A			Total
	Score 3	Score 4	Score 5	
Score 3	6	1	0	7 (58.3%)
Score 4	0	0	1	1 (8.3%)
Score 5	0	1	3	4 (33.3%)
Total	6 (50%)	2 (16.7%)	4 (33.3%)	12

*Weighted kappa k-value = 0.735

to decrease the radiation dose which inversely proportional to the beam pitch when other parameters are held constant.⁽²⁾ Moreover, this should be realized when beam pitch increases on tube current modulation technique, the tube current will automatically increase to preserve the noise conditions in the image. Thus the radiation dose does not decrease^(2,11)

Yoshiharu O, *et al*⁽¹²⁾ showed the radiation dose reduction 15% without significant deterioration of detection capability. The value was lesser approximately twice than this study even though they used higher beam pitch (0.83 and 1.48) than this study. They mentioned, if the image quality of their low dose protocol is recommended to be equal to the standard protocol, the radiation dose could be reduced to 50%. However, the CT number of nodule is -800 HU and -630 HU for image quality assessment were used in their study that difference from our study.

The average CNR value of high beam pitch technique (1.141) is lower than the standard beam pitch technique (0.798) because the high beam pitch provide high image noise due to the less photon reach the detectors. However, there is a large variation in average CNR value of small size nodules. The main variable factors result from the unstable sites of circular ROI on a very small size nodule. This is direct effect on variation of mean CT number of nodule and the standard deviation value of the background.

In clinical diagnostic, the lung nodules that have a clinical significance are non-calcified pulmonary nodule with 30 to 40 HU and ground glass opacities with -800 HU and -630 HU⁽¹²⁻¹⁴⁾ which lower than the CT number of simulated nodule with 100 HU of lung man chest phantom in this study that

presented the calcified nodules. This is the limitation of this study.

Although increasing beam pitch results in decreasing CNR, a small change of beam pitch still maintains the qualitative image quality. In other words, there is a little change in the intensity and contrast of the nodule. Thus it doesn't affect the nodule detectability. However, decreasing CNR owing to tube current reduction should be concerned since it could be affected the nodule detectability.⁽¹⁵⁾

The uncertainty of reviewer assessment has a little effect because the criteria for image scoring is clear and location of simulated nodule were set. The reviewers might have fatigue of the eyes.⁽¹⁶⁾ However, the inter-observer agreements are substantial good even though the beam pitch is changed when scanning. This shows that our results are reproducible between the two reviewers and reliable for detection of nodule with standard beam pitch and high beam pitch technique.^(9,12)

The benefit of this study is the influence of beam pitch related to the radiation dose and image quality which applied in protocol setting by the user according to the ALARA principle. Moreover, as the beam pitch increases, the scan time becomes shorter⁽²⁾ which is beneficial.

Conclusions

The results of this phantom study demonstrate that beam pitch and tube current is an important factor for CT chest scan. The effect of high beam pitch CT shows lower radiation dose and shorter scan time when compared to standard beam pitch with maintaining diagnostic image quality. Selection of the proper beam pitch results in the reduction of

patient's radiation dose. Although the optimization of chest CT protocol is excluded from this study, these recommendations should be applied for valuable in chest imaging in the future.

Acknowledgements

The authors would like to thank, Associate Professor Sivalee Suriyapee, and Mrs. Phetchareeya Suwanpradit for their time and effort. Lastly, the staff of the Computed Tomography Unit, King Chulalongkorn Memorial Hospital for their kind supports.

References

1. Kubo T, Ohno Y, Kauczor HU, Hatabu H. Radiation dose reduction in chest CT-review of available options. *Eur J Radiol* 2014 Oct; 83(1): 1953-61
2. Kubo T, Lin PJ, Stiller W, Takahashi M, Kauczor HU, Ohno Y, Hatabu H. Radiation Dose Reduction in Chest CT: a review. *Am J Radiol* 2008 Feb;190(2):335-43
3. Rehani MM. Radiation protection in newer imaging technologies. *Radiat Prot Dosimetry* 2010 Apr; 139(1 - 3): 357 - 62
4. Lee CH, Goo JM, Ye HJ, Ye SJ, Park CM, Chun EJ, Im JG. Radiation dose modulation techniques in the multidetector CT era: from basics to practice. *Radiographics* 2008 Sep; 28(5):1451-9
5. Lell MM, May M, Deak P, Alibek S, Kuefner M, Kuettner A, Köhler H, Achenbach S, Uder M, Radkowi T. High-pitch spiral computed tomography effect on image quality and radiation dose in pediatric chest computed tomography. *Invest Radiol* 2011 Feb;46(2): 116-23
6. McCollough CH, Primak AN, Braun N, Kofler J, Yu L, Christner J. Strategies for Reducing Radiation Dose in CT. *Radiol Clin North Am* 2009 Jan; 47(1): 27-40
7. Saini S. Multidetector row CT: principles and practice for abdominal applications. *Radiology* 2004 Nov;233(2):323-7
8. International Atomic Energy Agency. Quality Assurance program for computed tomography: diagnostic and therapy application. IAEA Human Health Series No 19. Vienna, Austria: IAEA, 2012:19-23
9. Svanholm H, Starklint H, Gundersen HJ, Fabricius J, Barlebo H, Olsen S. Reproducibility of histomorphologic diagnoses with special reference to the kappa statistic. *APMIS* 1989 Aug;97(8):689-98
10. Ulzheimer S, Flohr T. Multislice CT: Current Technology and Future Developments. In Reiser MF, Becker CR, Nikolaou K, Glazer G, eds. *Multislice CT*. London: Springer Berlin Heidelberg, 2009:3-23
11. Mahesh M, Scatarige JC, Cooper J, Fishman EK. Dose and pitch relationship for a particular multislice CT scanner. *AJR Am J Roentgenol* 2001 Dec;177(6):1273-5
12. Yoshiharu O, Hisanobu K, Astushi K, Mari T, Hiroyasu I, Sumiaki M, Kazuro S. Influence of detector collimation and beam pitch for identification and image quality of ground-glass attenuation and nodules on 16- and 64-detector row CT systems: Experimental study using chest phantom. *Eur J Radiol*

- 2007 Dec;64(3):406–13
13. Funama Y, Awai K, Liu D, Oda S, Yanaga Y, Nakaura T, Kawanaka K, Shimamura M, Yamashita Y. Detection of nodules showing ground-glass opacity in the lungs at low-dose multidetector computed tomography: phantom and clinical study. *J Comput Assist Tomogr* 2009 Jan;33(1):49-53
14. Matsumoto K, Ohno Y, Koyama H, Kono A, Inokawa H, Onishi Y, Nogami M, Takenaka D, Araki T, Sugimura K. 3D automatic exposure control for 64-detector row CT: radiation dose reduction in chest phantom study. *Eur J Radiol* 2011 Mar;77(3):522-27
15. Tang K, Wang L, Li R, Lin J, Zheng X, Cao G. Effect of Low Tube Voltage on Image Quality, Radiation Dose, and Low-Contrast Detectability at Abdominal Multidetector CT: Phantom Study. *J Biomed Biotechnol* 2012; 2012:13016
16. Robinson PJ. Radiology's Achilles' heel: error and variation in the interpretation of Roengen image. *Br J Radiol* 1997 Nov;70(839): 1085-98